Proof and Tralle hydrometer is to be used with distilled alcohol products with specific gravities less than 1.000.

Tralle 0 to 100% this scale will determine the direct percentage of alcohol in your spirits.

Proof 0 to 200 This is the US International Revenue scale for proof of spirits.

**FIRST**: Take the temperature of the sample to be tested.

TO USE YOUR HYDROMETER: Put a sample of liquid in a test jar. Gently lower the hydrometer into the liquid until it is floating. Spin the hydrometer to release air bubbles. At eye level, take a reading at the flat (not the meniscus) where the hydrometer emerges from the liquid.

Temperature reading below 60°F add to the observed hydrometer reading.

Temperature reading above 60°F substract from the observed hydrometer reading.

D	_			
Degrees		Proof 25 to	Proof 50	
Fahrenheit (F°)	under 25	under 50	and over	Tralle
100	14	12	16	8 7
95	12	10.5	14	
90	10	9	12	6
85	8.5	7.5	10	5
80	7	6	8	2
75	5	4.5	6	2 3 2
70	3	3	4	2
65	1.5	1.5	2	1
60	0	0	0	0
55	1.5	1.5	2	1
50	3.5	3	4	2 3 4
45	5	4.5	6	3
40	7	6	8	
35	9	8	10	5 6
30	10.5	9	12	6
25	12	10.5	14	7
20	14	12	16	8
15	16	13.5	18	9
10	18	15	20	10
5	19	16.5	22	11
0	21	18	24	12